

CTA Breast Reconstruction

Updated 07/19/25

Reviewed 07/19/25

Also Known As - inferior epigastric perforator, DIEP, DIEA, SIEP, SIEA protocol

Indications - surgical planning prior to breast reconstruction.

Bill under CT Angio Combo AP w/ and w/o Contrast Charge.

GENERAL SCAN NOTES

Move the patient's arms over his/her head if possible. Remove any metal from the imaging field of view.

Topogram - lung bases through pubic symphysis (obtained during end inspiration).

Craniocaudal scan coverage - lung bases through pubic symphysis (obtained during end inspiration).

Adjust FOV (field of view) on topogram to smallest without cropping anatomy.

IV Contrast:

Administer weight-based **Omnipaque-350** - **1 mL/kg** up to **150 mL** (100 mL minimum).

Inject at **4 mL/sec** followed by 40 mL saline flush, 20-gauge or larger in forearm or more proximal.

Bolus track off **distal abdominal aorta** triggered at **100 HU** + **4 secs**.

After bolus-tracking trigger, **scan patient caudocranial** (from pubic symphysis to lung bases).

Oral Contrast: generally not given for this protocol.

For **GE scanners**, it is essential for the 1st recon thickness on the scanner to match the 1st recon thickness in this protocol book for the prescribed Noise Index to be valid. The 1st recon should generally be the thickest recon in the protocol.

CTA Breast Reconstruction

SIEMENS PARAMETERS & RECONS

	Scan Mode	kV	mAs	Care Dose	Care kV & Lvl	Pitch	Acq	Coll	Rot Time	Scan Time
Sensation 16	spiral	120	150	on	NA	1.15	16	0.75	0.5	18.1
Go Up 32	spiral	110	92	on	on 115	1.50	32	0.7	0.8	11.9
Sensation 64	spiral	120	150	on	NA	1.20	64	0.6	0.5	10.9
Definition 64	spiral	120	120	on	on	1.40	64	0.6	0.5	9.3
Go Top 64	spiral	100	101	on	on 115	1.50	64	0.6	0.33	2.9
Drive 128	spiral	120	84	on	on	1.20	128	0.6	0.5	5.4
Force 192	spiral	110	90	on	on	1.20	192	0.6	0.5	3.6

AX ANGIO	3.0	3.0	Br40 / B41f	mediastinum	3	head/feet
AX ANGIO THINS	0.75	0.6	Br40 / B41f	mediastinum	3	head/feet
COR ANGIO	3.0	3.0	Br40 / B41f	mediastinum	3	front/back
SAG ANGIO	3.0	3.0	Br40 / B41f	mediastinum	3	left/right
AX MIPS	5.0	2.0	Br40 / B41f	mediastinum	3	head/feet
COR MIPS	5.0	2.0	Br40 / B41f	mediastinum	3	front/back
3D VRT (spin)	0.75	0.5	Bv36 / B31f	CT angio		

GE PARAMETERS & RECONS

	Scan Type	SFOV	kV	mA Range	Noise Index	Smart mA	Slice Thick	Beam Coll	Pitch	Speed	Rot Time	Dose Red	ASIR	Scan Time
LS 16	helical	large	120	100-440	16.36	on	2.5	20	1.375	27.50	0.5	NA	NA	9.1
Opt 540	helical	large	120	100-440	16.36	on	2.5	20	1.375	27.50	0.5	NA	NA	9.1
LS VCT 64	helical	large body	120	100-700	14.14	on	2.5	40	1.375	55.00	0.5	50	50	4.5
Disc VCT 64	helical	large body	120	100-700	14.14	on	2.5	40	1.375	55.00	0.5	50	50	4.5

AX ANGIO	2.5	2.5	std full	400/40	head/feet
COR ANGIO	2.5	2.5	std full	400/40	front/back
SAG ANGIO	2.5	2.5	std full	400/40	left/right
AX ANGIO THINS	0.75	0.6	std full	400/40	head/feet
AX MIPS	5.0	2.0	std full	400/40	head/feet
COR MIPS	5.0	2.0	std full	400/40	front/back
3D VRT (spin)	0.625	0.625	std full	400/40	

Must be first recon.

CTA Breast Reconstruction

PHILIPS PARAMETERS & RECONS

	Scan Mode	kV	Avg mAs	Dose Index	3D Dose	Pitch	Detect	Colli	Rot Time	Scan Time
Incisive 128	helical	120	103	20	on	1.00	64	0.625	0.75	9.4

AX ANGIO	3.0	3.0	B	mediastinum	3	head/feet
AX ANGIO THINS	0.75	0.6	B	mediastinum	3	head/feet
COR ANGIO	3.0	3.0	B	mediastinum	3	front/back
SAG ANGIO	3.0	3.0	B	mediastinum	3	left/right
AX MIPS	5.0	2.0	B	mediastinum	3	head/feet
COR MIPS	5.0	2.0	B	mediastinum	3	front/back
3D VRT (spin)	0.75	0.5				